

PRODUCT RECALL NOTICE

FLUKE 110V VOLTAGE TESTER T110

We have been made aware of some isolated incidents with this product where the voltage tester could display a false negative result, which can put the end user at risk of electric shock.

The product affected is:

Code	Product Description	Barcode	Model Number
2847J	FLUKE 110V VOLTAGE TESTER T110	095969599689	T110

These products were offered for sale between 06/03/2016 and 25/10/2018.



If you have purchased this product from us, please stop using it immediately.

In order to obtain a replacement product, please visit the Fluke recall website at <https://www.fluke.com/en-gb/support/safety-notice/tpt-recall> and follow the instructions given. The manufacturer will arrange collection of your product and send out a new and improved T110, free of charge.

Alternatively, please return it to your local **Screwfix Trade Counter** where you will be given a refund to the full value of your purchase. We can arrange collection of the item, please contact us on **03330 112 112** for a collection to be arranged and a full refund provided. Please visit our website for more information: www.screwfix.com.

If you have purchased the Fluke 110V Voltage Tester T110 for customers in the course of your own business, please contact them immediately, advise them to stop using the voltage tester and make arrangements for it to be removed.

Thank you in advance for your co-operation in this matter and we apologise for any inconvenience caused.